

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000966  
8:32 09/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:33  
STERILIZING 3:36  
136.0 C 228kPa 3:36  
136.3 C 230kPa 4:06  
135.9 C 229kPa 4:36  
136.4 C 230kPa 5:06  
136.4 C 232kPa 5:36  
136.2 C 228kPa 6:06  
136.4 C 233kPa 6:36  
136.5 C 232kPa 7:06  
136.5 C 232kPa 7:06  
VENTING 7:06  
AIR DRYING 7:59  
CYCLE COMPLETE 8:04  
-----

Digital Signature #  
AB6AAD9EFA9D4346  
-----