
Midmark M11 - v1.0.0
Total Cycles: 831

Name: RITTER 1
Sterilizer ID: V2178198
05/20/2021 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.4 F 0.0
2:00 174.4 F 3.9
4:00 196.8 F 9.2
6:00 228.4 F 6.8
8:00 226.7 F 10.6
10:00 232.6 F 7.3
12:00 241.8 F 11.7
14:00 244.9 F 12.6
16:00 258.2 F 19.7
18:00 268.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 29.0
1:00 273.0 F 29.7
2:00 272.0 F 28.9
3:00 272.7 F 29.7
4:00 271.6 F 29.0

Min 271.5 F 28.9
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 241.8 F 0.0
10:00 226.2 F 0.0
15:00 229.4 F 0.0
20:00 226.6 F 0.0
25:00 228.8 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

05/20/2021 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____