

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3695

Name: RITTER2  
Sterilizer ID: V2313228  
06/13/2024 12:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.8 F 0.0  
2:00 132.1 F 1.4  
4:00 146.4 F 4.0  
6:00 170.1 F 7.3  
8:00 190.5 F 10.9  
10:00 219.4 F 5.2  
12:00 225.5 F 7.8  
14:00 236.9 F 12.1  
16:00 232.8 F 8.5  
18:00 242.9 F 12.5  
20:00 243.4 F 12.3  
22:00 254.7 F 18.2  
24:00 265.2 F 24.6  
26:00 270.9 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 272.1 F 29.2  
2:00 272.9 F 29.6  
3:00 272.4 F 29.9  
4:00 271.5 F 29.3

Min 271.5 F 29.2  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.8 F 0.0  
5:00 217.6 F 0.0  
10:00 193.5 F 0.0  
15:00 191.1 F 0.0  
20:00 187.8 F 0.0  
25:00 185.8 F 0.0  
30:00 184.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 26:40  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:05:13  
CYCLE COMPLETE

06/13/2024 01:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_