
Midmark M11 - v1.0.5
Total Cycles: 1110

Name: RITTER2
Sterilizer ID: V2313228
05/18/2022 01:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.5 F 0.0
2:00 165.1 F 2.9
4:00 178.7 F 6.8
6:00 206.6 F 11.7
8:00 218.1 F 6.9
10:00 233.3 F 12.4
12:00 234.9 F 10.0
14:00 236.4 F 8.8
16:00 248.7 F 15.3
18:00 260.7 F 22.1
20:00 271.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.2 F 30.0
2:00 272.1 F 29.5
3:00 271.3 F 29.0
4:00 271.2 F 29.4

Min 271.0 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 218.2 F 0.0
10:00 229.1 F 0.0
15:00 231.7 F 0.0
20:00 228.8 F 0.0
25:00 227.6 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

05/18/2022 02:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____