
Midmark M11 - v1.0.0
Total Cycles: 1409

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2021 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.0 F 0.0
2:00 174.4 F 3.3
4:00 197.0 F 8.1
6:00 229.3 F 10.4
8:00 231.7 F 9.1
10:00 237.9 F 9.9
12:00 242.7 F 12.4
14:00 241.7 F 11.2
16:00 255.4 F 18.2
18:00 267.2 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.0
2:00 271.4 F 28.3
3:00 272.7 F 29.2
4:00 272.1 F 28.6

Min 271.4 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 230.5 F 0.0
10:00 214.9 F 0.0
15:00 218.4 F 0.0
20:00 219.5 F 0.0
25:00 220.0 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

11/30/2021 03:14 PM
ACCEPTED: X
REJECTED:
Operator: U02