
Midmark M11 - v1.0.0
Total Cycles: 2929

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2023 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.8 F 0.0
2:00 197.8 F 4.7
4:00 212.2 F 9.8
6:00 224.1 F 5.6
8:00 233.2 F 11.1
10:00 234.4 F 8.4
12:00 240.6 F 10.0
14:00 247.9 F 14.2
16:00 260.6 F 21.3
18:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.4 F 29.2
2:00 272.0 F 28.5
3:00 272.1 F 29.0
4:00 272.4 F 28.9

Min 271.6 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 234.0 F 0.0
10:00 237.3 F 0.0
15:00 230.2 F 0.0
20:00 225.6 F 0.0
25:00 224.3 F 0.0
30:00 224.2 F 0.0

Drying Complete

FILLING: 2:04
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:42
CYCLE COMPLETE

02/22/2023 12:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____