
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007442
7:48 28/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 1:58
STERILIZING 3:02
136.4 C 221kPa 3:02
136.4 C 221kPa 3:32
136.4 C 223kPa 4:02
136.4 C 223kPa 4:32
136.4 C 221kPa 5:02
136.4 C 220kPa 5:32
136.4 C 223kPa 6:02
136.4 C 223kPa 6:32
136.4 C 223kPa 6:32
VENTING 6:32
AIR DRYING 7:20
CYCLE COMPLETE 8:10

Digital Signature #
FD8036120F4CD6A8
