

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001400  
18:10 04/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:26  
PRESSURIZING 3:32  
STERILIZING 4:43  
136.0 C 221kPa 4:43  
136.3 C 224kPa 5:13  
136.4 C 223kPa 5:43  
136.2 C 223kPa 6:13  
136.3 C 224kPa 6:43  
136.4 C 223kPa 7:13  
136.3 C 224kPa 7:43  
136.4 C 225kPa 8:13  
136.4 C 225kPa 8:13  
VENTING 8:13  
AIR DRYING 9:08  
CYCLE COMPLETE 10:21  
-----

Digital Signature #  
3AB1391DFB810D80  
-----