
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002333
11:33 29/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:45
PRESSURIZING 1:50
STERILIZING 3:02
136.4 C 225kPa 3:02
136.4 C 226kPa 3:32
136.0 C 221kPa 4:02
136.4 C 223kPa 4:32
136.3 C 225kPa 5:02
136.4 C 226kPa 5:32
136.4 C 225kPa 6:02
136.4 C 226kPa 6:32
136.4 C 226kPa 6:32
VENTING 6:32
AIR DRYING 7:22
CYCLE COMPLETE 37:22

Digital Signature #
DEF568211DFCBA55
