
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001946
10:34 12/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:30
STERILIZING 3:39
136.0 C 222kPa 3:39
136.2 C 222kPa 4:09
136.2 C 224kPa 4:39
136.3 C 224kPa 5:09
136.3 C 223kPa 5:39
136.4 C 225kPa 6:09
136.4 C 226kPa 6:39
136.4 C 225kPa 7:09
136.4 C 225kPa 7:09
VENTING 7:09
AIR DRYING 8:05
CYCLE COMPLETE 8:41

Digital Signature #
4159C0A31AD38A71
