
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005406
11:13 21/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:25
STERILIZING 3:27
136.5 C 221kPa 3:27
136.1 C 223kPa 3:57
136.3 C 224kPa 4:27
136.4 C 225kPa 4:57
136.4 C 225kPa 5:27
136.3 C 223kPa 5:57
136.3 C 224kPa 6:27
136.2 C 223kPa 6:57
136.2 C 223kPa 6:57
VENTING 6:57
AIR DRYING 7:47
CYCLE COMPLETE 9:59

Digital Signature #
1BA2B571B34D549D
