
Midmark M11 - v1.0.0
Total Cycles: 3693

Name: RITTER 1
Sterilizer ID: V2178198
08/28/2023 03:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.8 F 0.0
2:00 178.0 F 3.3
4:00 189.2 F 7.3
6:00 206.3 F 11.8
8:00 223.5 F 6.1
10:00 229.6 F 10.9
12:00 231.9 F 7.2
14:00 244.8 F 13.2
16:00 242.8 F 11.7
18:00 254.8 F 17.7
20:00 265.3 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.7 F 29.0
2:00 271.2 F 28.3
3:00 272.6 F 29.1
4:00 271.6 F 28.4

Min 271.1 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 229.5 F 0.0
10:00 226.7 F 0.0
15:00 227.7 F 0.0
20:00 228.2 F 0.0
25:00 235.7 F 0.0
30:00 216.9 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:18
CYCLE COMPLETE

08/28/2023 04:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____