

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001906  
7:46 29/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:56  
PRESSURIZING 3:01  
STERILIZING 4:12  
136.2 C 222kPa 4:12  
136.2 C 221kPa 4:42  
136.3 C 224kPa 5:12  
136.2 C 222kPa 5:42  
136.4 C 224kPa 6:12  
136.2 C 221kPa 6:42  
136.2 C 223kPa 7:12  
136.4 C 224kPa 7:42  
136.4 C 224kPa 7:42  
VENTING 7:42  
AIR DRYING 8:31  
CYCLE COMPLETE 38:31

-----  
Digital Signature #  
BD73678F8EBE801E  
-----