
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000432
7:58 28/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:54
STERILIZING 4:00
136.0 C 230kPa 4:00
136.4 C 231kPa 4:30
136.4 C 233kPa 5:00
136.2 C 231kPa 5:30
136.4 C 232kPa 6:00
136.2 C 231kPa 6:30
136.2 C 231kPa 7:00
136.3 C 231kPa 7:30
136.3 C 231kPa 7:30
VENTING 7:30
AIR DRYING 8:22
CYCLE COMPLETE 8:53

Digital Signature #
171B4A6E8803CB1F
