
Midmark M11 - v1.0.5
Total Cycles: 3849

Name: RITTER2
Sterilizer ID: V2313228
08/12/2024 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.1 F 0.0
2:00 134.2 F 1.8
4:00 144.6 F 4.6
6:00 177.8 F 8.3
8:00 194.3 F 12.6
10:00 217.8 F 6.5
12:00 230.1 F 11.1
14:00 230.4 F 7.7
16:00 241.3 F 13.3
18:00 241.7 F 11.9
20:00 253.7 F 18.0
22:00 264.3 F 24.4
24:00 271.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.8
1:00 272.6 F 30.0
2:00 272.3 F 30.4
3:00 271.4 F 29.8
4:00 272.1 F 30.1

Min 271.4 F 29.7
Max 273.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 227.9 F 0.0
10:00 220.3 F 0.0
15:00 223.8 F 0.0
20:00 217.4 F 0.0
25:00 218.5 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:03:33
CYCLE COMPLETE

08/12/2024 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____