
Midmark M11 - v1.0.0
Total Cycles: 3269

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2023 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.3 F 0.0
2:00 143.2 F 0.0
4:00 160.1 F 0.5
6:00 189.0 F 4.5
8:00 209.5 F 9.1
10:00 226.3 F 11.0
12:00 231.7 F 8.5
14:00 243.1 F 13.3
16:00 240.2 F 10.7
18:00 236.2 F 8.7
20:00 250.6 F 15.3
22:00 261.9 F 21.7
24:00 272.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.6 F 28.7
2:00 271.7 F 28.1
3:00 271.6 F 28.4
4:00 272.5 F 28.7

Min 271.0 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 228.5 F 0.0
10:00 227.9 F 0.0
15:00 227.8 F 0.0
20:00 224.0 F 0.0
25:00 222.5 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:32
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:02:45
CYCLE COMPLETE

05/23/2023 02:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____