
Midmark M11 - v1.0.0
Total Cycles: 3152

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2023 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.4 F 0.0
2:00 175.3 F 0.0
4:00 202.2 F 0.7
6:00 216.7 F 5.8
8:00 228.9 F 10.6
10:00 233.2 F 7.2
12:00 243.4 F 12.9
14:00 241.7 F 11.3
16:00 240.5 F 10.4
18:00 254.9 F 17.7
20:00 267.0 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.5 F 28.5
2:00 271.6 F 28.2
3:00 272.5 F 28.6
4:00 271.5 F 28.0

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 211.4 F 0.0
10:00 226.5 F 0.0
15:00 231.6 F 0.0
20:00 228.0 F 0.0
25:00 222.2 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

04/24/2023 12:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____