
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005225
11:37 14/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:58
STERILIZING 2:48
136.0 C 220kPa 2:48
136.4 C 223kPa 3:18
136.4 C 224kPa 3:48
136.2 C 222kPa 4:18
136.4 C 225kPa 4:48
136.2 C 222kPa 5:18
136.4 C 219kPa 5:48
136.4 C 221kPa 6:18
136.4 C 221kPa 6:18
VENTING 6:18
AIR DRYING 8:09
CYCLE COMPLETE 11:15

Digital Signature #
54EE52F00D36C8CD
