

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1843

Name: RITTER 1  
Sterilizer ID: V2178198  
04/19/2022 01:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.9 F 0.0  
2:00 191.9 F 4.5  
4:00 207.9 F 10.0  
6:00 220.9 F 5.6  
8:00 232.2 F 10.9  
10:00 233.3 F 8.2  
12:00 240.2 F 10.3  
14:00 247.0 F 14.1  
16:00 259.9 F 21.2  
18:00 271.6 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.5 F 29.3  
2:00 271.4 F 28.7  
3:00 272.4 F 29.4  
4:00 271.8 F 28.8  
  
Min 271.3 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 223.7 F 0.0  
10:00 235.8 F 0.0  
15:00 230.4 F 0.0  
20:00 228.9 F 0.0  
25:00 229.6 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 18:36  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:44  
CYCLE COMPLETE

04/19/2022 02:49 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_