
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000752
10:12 27/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:09
STERILIZING 4:13
136.4 C 232kPa 4:13
136.2 C 231kPa 4:43
136.4 C 231kPa 5:13
136.4 C 233kPa 5:43
136.4 C 233kPa 6:13
136.3 C 233kPa 6:43
136.4 C 231kPa 7:13
136.4 C 233kPa 7:43
136.4 C 233kPa 7:43
VENTING 7:43
AIR DRYING 8:42
CYCLE COMPLETE 8:49

Digital Signature #
C06B96DBBEC9073E
