

-----  
Midmark M11 - v1.0.5  
Total Cycles: 200

Name: RITTER2  
Sterilizer ID: V2313228  
07/13/2021 10:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
2:00 199.7 F 3.8  
4:00 207.8 F 9.6  
6:00 222.6 F 7.3  
8:00 234.5 F 11.5  
10:00 235.9 F 9.8  
12:00 237.5 F 9.2  
14:00 250.7 F 16.0  
16:00 262.6 F 23.0  
18:00 272.7 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.1  
1:00 272.8 F 29.6  
2:00 271.8 F 29.3  
3:00 272.6 F 29.5  
4:00 271.7 F 29.2  
  
Min 271.6 F 28.9  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 219.8 F 0.0  
10:00 229.0 F 0.0  
15:00 228.8 F 0.0  
20:00 225.6 F 0.0  
25:00 225.0 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 2:31  
HEATING: 18:16  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:57  
CYCLE COMPLETE

07/13/2021 11:15 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01