
Midmark M11 - v1.0.5
Total Cycles: 3469

Name: RITTER2
Sterilizer ID: V2313228
03/22/2024 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 72.7 F 0.0
4:00 74.4 F 0.0
6:00 99.3 F 0.7
8:00 137.4 F 2.9
10:00 167.5 F 6.0
12:00 184.8 F 9.5
14:00 203.1 F 11.4
16:00 214.2 F 6.1
18:00 225.6 F 10.1
20:00 231.0 F 9.0
22:00 239.0 F 11.1
24:00 236.1 F 8.9
26:00 249.2 F 15.1
28:00 260.3 F 21.3
30:00 270.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.6 F 29.2
2:00 271.2 F 28.8
3:00 272.6 F 29.3
4:00 271.2 F 28.6

Min 271.0 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 230.0 F 0.0
10:00 210.3 F 0.0
15:00 184.5 F 0.0
20:00 186.0 F 0.0
25:00 188.1 F 0.0
30:00 188.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:51
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:08:59
CYCLE COMPLETE

03/22/2024 09:34 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____