
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000665
8:35 14/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:22
STERILIZING 4:34
136.3 C 230kPa 4:34
136.1 C 229kPa 5:04
136.4 C 232kPa 5:34
136.4 C 232kPa 6:04
136.3 C 230kPa 6:34
136.4 C 232kPa 7:04
136.2 C 230kPa 7:34
136.4 C 232kPa 8:04
136.4 C 232kPa 8:04
VENTING 8:04
AIR DRYING 8:59
CYCLE COMPLETE 9:03

Digital Signature #
565370E598E8F808
