

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001842  
8:31 15/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:14  
STERILIZING 4:20  
136.0 C 221kPa 4:20  
136.4 C 225kPa 4:50  
136.2 C 223kPa 5:20  
136.4 C 226kPa 5:50  
136.3 C 224kPa 6:20  
136.4 C 226kPa 6:50  
136.4 C 226kPa 7:20  
136.4 C 225kPa 7:50  
136.4 C 225kPa 7:50  
VENTING 7:50  
AIR DRYING 8:41  
CYCLE COMPLETE 8:44  
-----

Digital Signature #  
001B68DAD91CDE0B  
-----