

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4147

Name: RITTER 1  
Sterilizer ID: V2178198  
12/21/2023 11:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.1 F 0.0  
2:00 141.9 F 1.5  
4:00 144.2 F 3.9  
6:00 180.9 F 7.3  
8:00 205.2 F 11.8  
10:00 211.2 F 4.0  
12:00 221.6 F 8.0  
14:00 233.6 F 12.5  
16:00 234.5 F 9.0  
18:00 241.7 F 11.1  
20:00 244.4 F 12.5  
22:00 255.8 F 18.4  
24:00 266.0 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.4 F 29.0  
2:00 272.1 F 28.6  
3:00 271.1 F 28.2  
4:00 272.0 F 29.0

Min 271.1 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 229.1 F 0.0  
10:00 235.5 F 0.0  
15:00 226.8 F 0.0  
20:00 225.7 F 0.0  
25:00 230.3 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 25:47  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:04:35  
CYCLE COMPLETE

12/21/2023 12:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_