
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005986
12:24 17/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:10
STERILIZING 4:12
136.4 C 223kPa 4:12
136.2 C 223kPa 4:43
136.1 C 222kPa 5:13
136.4 C 225kPa 5:43
136.2 C 221kPa 6:13
136.4 C 225kPa 6:43
136.4 C 224kPa 7:13
136.4 C 224kPa 7:43
136.4 C 224kPa 7:43
VENTING 7:43
AIR DRYING 8:34
CYCLE COMPLETE 8:42

Digital Signature #
9E4FCF0DBBF34D85
