
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000859
15:15 12/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 3:02
STERILIZING 4:12
136.0 C 222kPa 4:12
136.4 C 229kPa 4:42
136.4 C 231kPa 5:12
136.2 C 229kPa 5:42
136.4 C 233kPa 6:12
136.1 C 229kPa 6:42
136.4 C 232kPa 7:12
136.3 C 231kPa 7:42
136.3 C 231kPa 7:42
VENTING 7:42
AIR DRYING 8:34
CYCLE COMPLETE 8:41

Digital Signature #
19A775B2932BFA2D
