
Midmark M11 - v1.0.0
Total Cycles: 1076

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2021 06:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 73.8 F 0.0
4:00 102.1 F 0.0
6:00 109.9 F 0.1
8:00 130.8 F 0.1
10:00 163.3 F 0.2
12:00 185.0 F 0.2
14:00 200.4 F 2.4
16:00 213.8 F 4.9
18:00 221.7 F 7.8
20:00 229.9 F 11.0
22:00 233.1 F 7.8
24:00 239.0 F 10.8
26:00 236.7 F 8.4
28:00 244.8 F 11.9
30:00 245.9 F 12.9
32:00 257.1 F 18.8
34:00 267.1 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.8 F 28.7
2:00 271.6 F 27.8
3:00 272.6 F 28.7
4:00 272.3 F 28.3

Min 271.5 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 199.8 F 0.0
10:00 204.7 F 0.0
15:00 210.8 F 0.0
20:00 201.4 F 0.0
25:00 197.9 F 0.0
Dry Cycle Aborted

FILLING: 0:49
HEATING: 35:31
STERILIZING: 4:00
VENTING: 3:21
DRYING: 29:41
TOTAL CYCLE: 01:13:22

08/09/2021 08:02 AM
ACCEPTED: X
REJECTED:
Operator: U01