
Midmark M11 - v1.0.5
Total Cycles: 319

Name: RITTER2
Sterilizer ID: V2313228
08/30/2021 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.2 F 0.0
2:00 162.3 F 3.0
4:00 176.9 F 7.4
6:00 202.4 F 12.7
8:00 217.3 F 7.2
10:00 232.0 F 12.5
12:00 236.4 F 10.5
14:00 235.8 F 9.3
16:00 250.1 F 16.2
18:00 262.2 F 23.0
20:00 272.6 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.9 F 29.5
2:00 272.4 F 30.0
3:00 271.7 F 29.3
4:00 272.3 F 29.8

Min 271.4 F 29.2
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 220.8 F 0.0
10:00 209.6 F 0.0
15:00 218.3 F 0.0
20:00 213.5 F 0.0
25:00 220.1 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

08/30/2021 10:43 AM
ACCEPTED: X
REJECTED:
Operator: U02