
Midmark M11 - v1.0.0
Total Cycles: 5292

Name: RITTER 1
Sterilizer ID: V2178198
01/29/2025 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 170.1 F 2.4
4:00 179.9 F 5.9
6:00 200.7 F 9.7
8:00 211.8 F 10.5
10:00 223.9 F 7.5
12:00 236.8 F 12.2
14:00 235.0 F 8.8
16:00 241.7 F 10.9
18:00 245.6 F 13.1
20:00 257.1 F 19.4
22:00 267.5 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 271.4 F 29.0
2:00 272.8 F 29.6
3:00 272.4 F 29.2
4:00 271.8 F 29.2

Min 271.0 F 28.5
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 219.9 F 0.0
10:00 235.1 F 0.0
15:00 243.3 F 0.0
20:00 222.3 F 0.0
25:00 215.1 F 0.0
30:00 209.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:41
STERILIZING: 4:00
VENTING: 4:07
DRYING: 30:00
TOTAL CYCLE: 01:02:52
CYCLE COMPLETE

01/29/2025 01:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____