
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005733
9:00 12/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:58
STERILIZING 3:06
136.4 C 221kPa 3:06
136.1 C 219kPa 3:36
136.0 C 218kPa 4:06
136.4 C 222kPa 4:36
136.4 C 223kPa 5:06
136.4 C 224kPa 5:36
136.4 C 224kPa 6:06
136.4 C 224kPa 6:36
136.4 C 224kPa 6:36
VENTING 6:37
AIR DRYING 7:22
CYCLE COMPLETE 7:40

Digital Signature #
3D1F7B68C3046284
