
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000539
12:17 25/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:24
PRESSURIZING 1:32
STERILIZING 2:41
136.0 C 230kPa 2:41
136.4 C 233kPa 3:11
136.2 C 232kPa 3:41
136.1 C 232kPa 4:11
136.4 C 233kPa 4:41
136.1 C 230kPa 5:11
136.4 C 233kPa 5:41
136.4 C 231kPa 6:11
136.4 C 231kPa 6:11
VENTING 6:11
AIR DRYING 7:08
CYCLE COMPLETE 22:08

Digital Signature #
9A14CC37DDAB96FC
