

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1111

Name: RITTER2  
Sterilizer ID: V2313228  
05/18/2022 02:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
2:00 183.3 F 3.9  
4:00 198.6 F 9.1  
6:00 218.4 F 7.2  
8:00 231.7 F 10.3  
10:00 233.3 F 8.0  
12:00 240.6 F 10.8  
14:00 247.8 F 14.5  
16:00 261.0 F 21.8  
18:00 272.5 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.5 F 29.2  
2:00 271.7 F 29.0  
3:00 272.6 F 29.3  
4:00 271.5 F 28.7  
  
Min 271.5 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.1 F 0.0  
5:00 221.2 F 0.0  
10:00 231.5 F 0.0  
15:00 227.1 F 0.0  
20:00 221.9 F 0.0  
25:00 219.6 F 0.0  
30:00 217.7 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 18:25  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 00:57:30  
CYCLE COMPLETE

05/18/2022 03:35 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_