
Midmark M11 - v1.0.0
Total Cycles: 1795

Name: RITTER 1
Sterilizer ID: V2178198
04/05/2022 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.4 F 0.0
2:00 143.3 F 0.9
4:00 161.7 F 4.2
6:00 182.7 F 8.6
8:00 207.2 F 11.9
10:00 223.4 F 7.0
12:00 234.8 F 11.9
14:00 234.1 F 8.4
16:00 241.7 F 11.1
18:00 244.7 F 12.7
20:00 257.1 F 19.2
22:00 268.2 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.8 F 29.2
2:00 271.6 F 28.4
3:00 272.2 F 29.1
4:00 272.5 F 28.9

Min 271.2 F 28.2
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 236.1 F 0.0
10:00 200.0 F 0.0
15:00 209.1 F 0.0
20:00 233.9 F 0.0
25:00 232.1 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:17
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

04/05/2022 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____