

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000036  
8:58 15/01/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:02  
PRESSURIZING 4:02  
STERILIZING 5:03  
136.4 C 223kPa 5:03  
136.3 C 223kPa 5:33  
136.4 C 222kPa 6:03  
136.3 C 223kPa 6:33  
136.4 C 220kPa 7:03  
136.4 C 221kPa 7:33  
136.4 C 224kPa 8:03  
136.4 C 221kPa 8:33  
136.4 C 221kPa 8:33  
VENTING 8:33  
AIR DRYING 9:27  
CYCLE COMPLETE 9:52

-----  
Digital Signature #  
F62366DADB39FE69  
-----