

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000322  
11:10 17/08/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:31  
PRESSURIZING 3:33  
STERILIZING 4:35  
136.0 C 220kPa 4:35  
136.4 C 224kPa 5:05  
136.4 C 222kPa 5:35  
136.1 C 222kPa 6:05  
136.4 C 224kPa 6:35  
136.4 C 224kPa 7:05  
136.3 C 223kPa 7:35  
136.3 C 224kPa 8:05  
136.3 C 224kPa 8:05  
VENTING 8:05  
AIR DRYING 9:00  
CYCLE COMPLETE 9:47  
-----

Digital Signature #  
4F44AFE5E5D1029D  
-----