
Midmark M11 - v1.0.5
Total Cycles: 4262

Name: RITTER2
Sterilizer ID: V2313228
01/14/2025 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 72.4 F 0.0
4:00 73.8 F 0.0
6:00 97.5 F 0.0
8:00 138.4 F 0.1
10:00 168.1 F 0.3
12:00 189.7 F 0.8
14:00 204.2 F 3.8
16:00 220.8 F 7.2
18:00 230.6 F 10.9
20:00 231.7 F 7.7
22:00 239.0 F 11.3
24:00 235.4 F 9.0
26:00 241.0 F 10.5
28:00 247.3 F 14.1
30:00 258.5 F 20.3
32:00 268.6 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.4
2:00 271.5 F 28.7
3:00 272.5 F 29.3
4:00 271.5 F 28.6

Min 271.4 F 28.5
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.4 F 0.0
5:00 224.6 F 0.0
10:00 228.8 F 0.0
15:00 230.7 F 0.0
20:00 228.6 F 0.0
25:00 225.0 F 0.0
30:00 223.7 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:14
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:11:39
CYCLE COMPLETE

01/14/2025 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____