

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002185  
13:43 29/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:03  
STERILIZING 3:03  
136.4 C 224kPa 3:03  
136.4 C 224kPa 3:33  
136.2 C 223kPa 4:03  
136.2 C 222kPa 4:33  
136.4 C 226kPa 5:03  
136.4 C 226kPa 5:33  
136.2 C 225kPa 6:03  
136.4 C 225kPa 6:33  
136.4 C 225kPa 6:33  
VENTING 6:33  
AIR DRYING 7:25  
CYCLE COMPLETE 37:25  
-----

Digital Signature #  
4D87561EB3DAF638  
-----