

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3248

Name: RITTER2  
Sterilizer ID: V2313228  
12/22/2023 09:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.0 F 0.0  
2:00 74.6 F 0.0  
4:00 86.4 F 0.4  
6:00 123.1 F 2.0  
8:00 158.0 F 5.2  
10:00 180.9 F 8.5  
12:00 206.9 F 12.8  
14:00 219.2 F 5.7  
16:00 228.7 F 9.8  
18:00 237.1 F 9.6  
20:00 239.5 F 10.7  
22:00 235.4 F 8.1  
24:00 248.8 F 14.7  
26:00 260.3 F 21.2  
28:00 270.9 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.5 F 28.9  
1:00 272.5 F 29.3  
2:00 272.6 F 29.5  
3:00 271.7 F 29.1  
4:00 272.1 F 28.9

Min 271.5 F 28.7  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
5:00 220.8 F 0.0  
10:00 204.2 F 0.0  
15:00 212.8 F 0.0  
20:00 208.6 F 0.0  
25:00 216.1 F 0.0  
30:00 215.1 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:19  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:07:15  
CYCLE COMPLETE

12/22/2023 10:35 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_