

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3549

Name: RITTER2  
Sterilizer ID: V2313228  
04/23/2024 10:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.2 F 0.0  
2:00 116.1 F 0.6  
4:00 126.8 F 2.1  
6:00 129.3 F 4.8  
8:00 158.4 F 8.4  
10:00 177.1 F 12.1  
12:00 209.5 F 3.9  
14:00 220.1 F 7.7  
16:00 233.7 F 11.9  
18:00 232.6 F 8.4  
20:00 243.7 F 12.7  
22:00 243.0 F 12.3  
24:00 254.2 F 18.1  
26:00 264.6 F 24.4  
28:00 271.0 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.6  
1:00 272.0 F 29.5  
2:00 272.1 F 29.5  
3:00 272.3 F 29.6  
4:00 272.5 F 29.8

Min 271.8 F 29.5  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
5:00 230.2 F 0.0  
10:00 230.2 F 0.0  
15:00 225.5 F 0.0  
20:00 229.1 F 0.0  
25:00 232.4 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 28:40  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:07:19  
CYCLE COMPLETE

04/23/2024 11:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_