
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000484
10:40 18/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:26
STERILIZING 4:26
136.4 C 223kPa 4:26
136.4 C 225kPa 4:56
136.4 C 223kPa 5:26
136.4 C 224kPa 5:56
136.3 C 224kPa 6:26
136.1 C 222kPa 6:56
136.4 C 225kPa 7:26
136.4 C 224kPa 7:56
136.4 C 224kPa 7:56
VENTING 7:56
AIR DRYING 8:51
CYCLE COMPLETE 9:26

Digital Signature #
943C9E218F1FDDE6
