

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1099

Name: RITTER2  
Sterilizer ID: V2313228  
05/11/2022 04:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
2:00 205.8 F 5.2  
4:00 227.2 F 12.0  
6:00 233.1 F 9.4  
8:00 236.0 F 9.0  
10:00 240.7 F 10.5  
12:00 249.4 F 15.2  
14:00 262.3 F 22.5  
16:00 272.4 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.3 F 29.0  
2:00 271.8 F 29.0  
3:00 272.2 F 29.0  
4:00 271.7 F 28.9  
  
Min 271.5 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
5:00 221.7 F 0.0  
10:00 233.7 F 0.0  
15:00 229.2 F 0.0  
20:00 224.7 F 0.0  
25:00 224.2 F 0.0  
30:00 223.8 F 0.0  
Drying Complete

FILLING: 3:45  
HEATING: 16:13  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:57:04  
CYCLE COMPLETE

05/12/2022 06:52 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02