
Midmark M11 - v1.0.5
Total Cycles: 2559

Name: RITTER2
Sterilizer ID: V2313228
06/23/2023 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 176.5 F 1.8
4:00 182.6 F 6.0
6:00 204.9 F 11.4
8:00 222.6 F 6.8
10:00 235.6 F 12.4
12:00 237.2 F 10.3
14:00 235.6 F 8.8
16:00 250.4 F 16.0
18:00 262.3 F 22.9
20:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.9 F 29.8
2:00 271.8 F 29.1
3:00 271.9 F 29.7
4:00 272.4 F 29.7

Min 271.4 F 28.9
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 222.7 F 0.0
10:00 233.8 F 0.0
15:00 222.3 F 0.0
20:00 234.0 F 0.0
25:00 230.3 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

06/23/2023 01:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____