
Midmark M11 - v1.0.5
Total Cycles: 4202

Name: RITTER2
Sterilizer ID: V2313228
12/11/2024 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 72.9 F 0.0
4:00 82.0 F 0.0
6:00 119.7 F 0.1
8:00 160.0 F 1.3
10:00 178.8 F 4.7
12:00 194.7 F 8.2
14:00 208.0 F 12.0
16:00 218.0 F 5.4
18:00 228.6 F 9.5
20:00 236.4 F 10.8
22:00 240.0 F 11.5
24:00 238.4 F 10.1
26:00 251.6 F 16.6
28:00 263.0 F 23.2
30:00 272.4 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.5 F 29.3
2:00 272.2 F 29.5
3:00 272.1 F 29.0
4:00 272.6 F 29.7

Min 271.6 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.8 F 0.0
5:00 206.5 F 0.0
10:00 220.6 F 0.0
15:00 223.9 F 0.0
20:00 218.9 F 0.0
25:00 217.0 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:13
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:08:25
CYCLE COMPLETE

12/11/2024 10:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____