
Midmark M11 - v1.0.0
Total Cycles: 838

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2021 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.1 F 0.0
2:00 157.2 F 2.0
4:00 173.3 F 5.4
6:00 199.9 F 9.6
8:00 232.2 F 7.1
10:00 225.2 F 9.0
12:00 241.0 F 10.8
14:00 242.4 F 11.6
16:00 239.3 F 9.9
18:00 252.8 F 16.6
20:00 264.4 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.7 F 28.8
2:00 272.3 F 28.7
3:00 272.1 F 28.6
4:00 272.1 F 28.6

Min 270.7 F 28.0
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.6 F 0.0
5:00 244.9 F 0.0
10:00 233.7 F 0.0
15:00 230.5 F 0.0
20:00 226.5 F 0.0
25:00 229.5 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:49
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:23
CYCLE COMPLETE

05/26/2021 01:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____