
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001660
13:16 31/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:47
STERILIZING 4:49
136.4 C 223kPa 4:49
136.2 C 224kPa 5:19
136.4 C 225kPa 5:49
136.2 C 223kPa 6:19
136.2 C 224kPa 6:49
136.4 C 222kPa 7:19
136.2 C 224kPa 7:49
136.3 C 222kPa 8:19
136.3 C 222kPa 8:19
VENTING 8:20
AIR DRYING 9:13
CYCLE COMPLETE 9:17

Digital Signature #
83C9A2E37813AEFE
