
Midmark M11 - v1.0.5
Total Cycles: 3114

Name: RITTER2
Sterilizer ID: V2313228
11/13/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 68.6 F 0.0
4:00 73.6 F 0.2
6:00 91.8 F 1.2
8:00 134.4 F 3.7
10:00 157.7 F 7.2
12:00 177.3 F 10.8
14:00 208.2 F 5.7
16:00 215.4 F 7.3
18:00 225.4 F 11.4
20:00 227.1 F 7.1
22:00 239.9 F 12.4
24:00 238.7 F 10.5
26:00 251.2 F 16.6
28:00 262.1 F 22.9
30:00 272.3 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.7 F 29.9
2:00 271.5 F 29.3
3:00 272.5 F 29.9
4:00 271.3 F 29.2

Min 271.3 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 205.7 F 0.0
10:00 200.4 F 0.0
15:00 200.7 F 0.0
20:00 206.9 F 0.0
25:00 215.2 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:28
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:08:34
CYCLE COMPLETE

11/13/2023 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____