

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1139

Name: RITTER2  
Sterilizer ID: V2313228  
05/30/2022 02:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.4 F 0.0  
2:00 132.6 F 1.5  
4:00 151.2 F 4.7  
6:00 176.9 F 9.0  
8:00 200.5 F 10.0  
10:00 217.7 F 7.9  
12:00 232.0 F 13.0  
14:00 235.3 F 10.0  
16:00 235.9 F 8.8  
18:00 247.9 F 14.9  
20:00 259.7 F 21.5  
22:00 270.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.5 F 29.9  
2:00 271.5 F 29.2  
3:00 272.1 F 30.0  
4:00 272.4 F 29.7

Min 271.2 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 204.7 F 0.0  
10:00 218.0 F 0.0  
15:00 224.4 F 0.0  
20:00 226.6 F 0.0  
25:00 225.2 F 0.0  
30:00 221.8 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:49  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:01:00  
CYCLE COMPLETE

05/30/2022 03:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_