
Midmark M11 - v1.0.5
Total Cycles: 3001

Name: RITTER2
Sterilizer ID: V2313228
10/12/2023 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.4 F 0.0
2:00 160.2 F 2.8
4:00 179.2 F 6.9
6:00 202.2 F 12.3
8:00 216.5 F 7.0
10:00 231.0 F 12.4
12:00 235.3 F 9.7
14:00 236.6 F 9.1
16:00 249.2 F 15.2
18:00 261.6 F 22.4
20:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.5 F 29.7
2:00 271.3 F 28.9
3:00 272.2 F 29.8
4:00 272.1 F 29.4

Min 271.0 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 241.9 F 0.0
10:00 225.4 F 0.0
15:00 225.5 F 0.0
20:00 227.2 F 0.0
25:00 231.2 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

10/12/2023 11:53 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____