
Midmark M11 - v1.0.0
Total Cycles: 4784

Name: RITTER 1
Sterilizer ID: V2178198
08/02/2024 12:23 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.1 F 0.0
2:00 125.6 F 0.1
4:00 160.3 F 0.4
6:00 187.2 F 1.1
8:00 210.3 F 5.7
10:00 221.5 F 9.4
12:00 232.4 F 12.3
14:00 231.6 F 8.1
16:00 243.9 F 13.5
18:00 240.6 F 10.8
20:00 236.0 F 8.6
22:00 253.2 F 16.8
24:00 266.4 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.6 F 28.9
2:00 272.3 F 28.2
3:00 271.8 F 28.3

Min 271.6 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 156.0 F 0.0
5:00 237.5 F 0.0
10:00 220.5 F 0.0
15:00 220.4 F 0.0
20:00 213.8 F 0.0
25:00 213.1 F 0.0
Dry Cycle Aborted

FILLING: 0:36
HEATING: 25:17
STERILIZING: 3:00
VENTING: 2:31
DRYING: 28:47
TOTAL CYCLE: 01:00:11

08/02/2024 01:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____